

Preface

30

31

Additional Copies

33

34 Additional copies are available from the Internet. You may also send an email request to [CDRH-](mailto:CDRH-Guidance@fda.hhs.gov)
35 [Guidance@fda.hhs.gov](mailto:CDRH-Guidance@fda.hhs.gov) to receive a copy of the guidance. Please include the document number
36 GUI00021016 and complete title of the guidance in the request.

37

DRAFT

Table of Contents

38		
39		
40	I. Introduction.....	1
41	II. Background.....	2
42	III. Scope.....	3
43	IV. Clinical Performance Testing Considerations	4
44	A. Study Design.....	5
45	B. Study Duration and Follow-up Schedule.....	6
46	C. Inclusion/Exclusion Criteria	6
47	D. Patient Demographics	7
48	E. Treatment Parameters/Protocol.....	8
49	F. Safety Endpoints and Data.....	8
50	G. Effectiveness Endpoints and Data	9
51	H. Adverse Events	10
52	I. Statistical Analysis Considerations.....	12
53	(1) Sample Size.....	12
54	(2) Analysis Methods.....	12
55	(3) Missing Data	13
56	(4) Subgroup Analyses	14
57	J. Pediatric Studies.....	14
58	V. Benefit-Risk Considerations	19
59	A. Benefit-Risk	19
60	B. Use of Modified Clavien-Dindo to Assess Risk.....	21
61	C. Balancing Weight Loss and Adverse Events for an Indication of Weight Loss.....	22
62		
63		

64 **Medical Devices with Indications**
65 **Associated with Weight Loss - Clinical**
66 **Study and Benefit-Risk Considerations**
67

68 **Draft Guidance for Industry and**
69 **Food and Drug Administration Staff**
70

71 *This draft guidance, when finalized, will represent the current thinking of the Food and Drug*
72 *Administration (FDA or Agency) on this topic. It does not establish any rights for any person*
73 *and is not binding on FDA or the public. You can use an alternative approach if it satisfies*
74 *the requirements of the applicable statutes and regulations. To discuss an alternative*
75 *approach, contact the FDA staff or Office responsible for this guidance as listed on the title*
76 *page.*

77
78 **I. Introduction**

79 This draft guidance document provides recommendations regarding clinical study design for
80 devices with indications for use associated with weight loss, and also includes discussion on how
81 FDA considers the benefit-risk analysis to support such indications.¹ Examples of indications
82 associated with weight loss include indications for weight loss, weight reduction, weight
83 management, or obesity treatment in patients who are overweight or have obesity.² Due to the
84 wide variety of device designs, among other things, there can be variability in the demonstrated
85 weight loss and risk associated with these devices. The recommendations reflect current review
86 practices of premarket submissions (e.g., Premarket Approval (PMA) Applications,
87 Investigational Device Exemption (IDE) Applications, Premarket Notifications (510(k)s), and
88 De Novo classification requests) for these devices and are intended to promote consistency and
89 facilitate efficient review of these submissions.

90
91 In general, FDA’s guidance documents do not establish legally enforceable responsibilities.
92 Instead, guidances describe the Agency’s current thinking on a topic and should be viewed only
93 as recommendations, unless specific regulatory or statutory requirements are cited. The use of

¹ For further information on how FDA considers benefit-risk factors when evaluating substantial equivalence in 510(k)s generally, see <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/benefit-risk-factors-consider-when-determining-substantial-equivalence-premarket-notifications-510k>.

² For further information on weight-loss and weight-management devices, see also <https://www.fda.gov/medical-devices/products-and-medical-procedures/weight-loss-and-weight-management-devices>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

94 the word *should* in Agency guidances means that something is suggested or recommended, but
95 not required.
96

97 **II. Background**

98 Prior to issuing this draft guidance, FDA requested public comment on a concept for balancing
99 the benefit of weight loss with the risks of adverse events in a discussion paper (September
100 2019).³ FDA considered public comments and incorporated the feedback as appropriate in
101 developing this draft guidance. The discussion paper continued FDA’s efforts to be transparent
102 and informative about how we regulate devices with indications associated with weight loss.
103

104 Additionally, FDA has previously engaged stakeholders regarding how we can help to ensure
105 patients have access to appropriately safe and effective devices indicated for weight loss:⁴

- 106 • On November 16-17, 2005, FDA’s Pediatric Advisory Committee held a public
107 meeting on Clinical Trial Design Issues for Pediatric Obesity Devices.⁵
- 108 • On October 16-18, 2011, FDA, Dartmouth Device Development/GI at Dartmouth
109 Medical School, and the Obesity, Metabolism and Nutrition Institute at Massachusetts
110 General Hospital co-sponsored a two-day workshop, “Device Development in
111 Obesity and Metabolic Disease (DDOMD).”
- 112 • On May 10-11, 2012, the Gastroenterology-Urology Devices Panel of the Medical
113 Devices Advisory Committee discussed general issues related to obesity treatment
114 devices and provided clinical study design recommendations to better evaluate the
115 safety and effectiveness of obesity treatment devices.⁶
- 116 • In 2013, FDA published a benefit-risk assessment paradigm that could provide an a
117 priori tool for systematic assessment of the risks associated with the devices intended
118 for treatment of obesity and to suggest appropriate levels of benefit for devices with
119 different risk levels.⁷

³ See Docket No. FDA-2019-N-4060 (<https://www.regulations.gov/docket?D=FDA-2019-N-4060>).

⁴ See also <https://www.fda.gov/medical-devices/weight-loss-and-weight-management-devices/fda-activities-weight-loss-and-weight-management-devices>.

⁵ FDA Pediatric Advisory Committee, *Development of Trials to Assess the Safety and Efficacy Relevant to Scientific and Ethical Issues Surrounding Trials for Pediatric Devices for Weight Loss*. Gaithersburg, MD. Meeting materials can be accessed at <https://wayback.archive-it.org/7993/2017040322257/https://www.fda.gov/ohrms/dockets/ac/oc05.html#Pediatric>.

⁶ 2012 Materials of the Gastroenterology-Urology Devices Panel can be accessed at <https://wayback.archive-it.org/7993/20170113191551/http://www.fda.gov/AdvisoryCommittees/CommitteesMeetingMaterials/MedicalDevices/MedicalDevicesAdvisoryCommittee/Gastroenterology-UrologyDevicesPanel/ucm286235.htm>.

⁷ Lerner, H., Whang, J., & Nipper, R. (2013). Benefit-risk paradigm for clinical trial design of obesity devices: FDA proposal. *Surgical endoscopy*, 27(3), 702-707.

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 120
- 121
- 122
- In 2015, FDA worked with the Research Triangle Institute Health Solutions (RTI-HS) to conduct the first national benefit-risk patient preference study to provide information on patient risk tolerance for weight loss devices.⁸
 - On June 28, 2018, FDA held a listening session with patients who have used FDA-approved devices with indications associated with weight loss.
- 123
- 124

125

126 FDA refers the reader to the Q-Submission Program throughout this guidance document. For

127 details on the Q-Submission Program, refer to the guidance “[Requests for Feedback and](#)

128 [Meetings for Medical Device Submissions: The Q-Submission Program](#).”⁹

129

130 **III. Scope**

131 The scope of this document is limited to devices with indications for use associated with weight

132 loss, including weight loss, weight reduction, weight management, or obesity treatment in

133 patients who are overweight or have obesity. This includes the existing product codes listed in

134 Table 1 below:

135

136 **Table 1. Existing product codes within the scope of this guidance**

Product Code	Product Code Name	Regulation Number
LTI	Intragastric implant for morbid obesity	Not applicable ¹⁰
OYF	Aspiration therapy system	Not applicable ¹¹
PIM	Neuromodulator for obesity	Not applicable ¹²
ONY	Oral removable retainer for weight management	21 CFR 876.5981 ¹³
QFQ	Ingested, Transient, Space Occupying Device For Weight Management And/Or Weight Loss	21 CFR 876.5982 ¹⁴
QTD	Endoscopic Suturing Device For Altering Gastric Anatomy For Weight Loss	21 CFR 876.5983 ¹⁵

⁸ Ho, M. P., Gonzalez, J. M., Lerner, H. P., Neuland, C. Y., Whang, J. M., McMurry-Heath, M., Hauber, A.B. & Irony, T. (2015). Incorporating patient-preference evidence into regulatory decision making. *Surgical endoscopy*, 29(10), 2984-2993.

⁹ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/requests-feedback-and-meetings-medical-device-submissions-q-submission-program>.

¹⁰ This is a postamendments class III device.

¹¹ *Ibid.*

¹² *Ibid.*

¹³ This classification regulation includes special controls. *See* 21 CFR 876.5981(b).

¹⁴ This classification regulation includes special controls. *See* 21 CFR 876.5982(b).

¹⁵ This classification regulation includes special controls established in the reclassification order, available at https://www.accessdata.fda.gov/cdrh_docs/pdf21/DEN210045.pdf. The publication of this classification in the Federal Register and codification in the Code of Federal Regulations are currently pending.

Contains Nonbinding Recommendations

Draft – Not for Implementation

137
138 Although the product codes listed above are current as of the date of issuance of this draft
139 guidance, new product codes or classification regulations may be created over time and could
140 fall within the scope of this guidance. We recommend that you reference the product code
141 database (<http://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfPCD/classification.cfm>) or
142 contact OHT3: Office of Gastro-Renal, ObGyn, General Hospital, and Urology Devices if you
143 are uncertain whether this guidance applies to your device and the product code for your device
144 is not already captured in this guidance.

145
146 Some of the recommendations in this guidance may assist in complying with some of the
147 special controls for devices with indications associated with weight loss. For information
148 regarding special controls for oral removable retainers for weight management, see 21 CFR
149 876.5981(b). For information regarding special controls for ingested, transient, space
150 occupying devices for weight management and/or weight loss, see 21 CFR 876.5982(b). For
151 information regarding special controls for endoscopic suturing devices for altering gastric
152 anatomy for weight loss, see FDA’s website.¹⁶

153
154 This draft guidance should be viewed as a complement to FDA’s draft guidance entitled,
155 “[Medical Devices with Indications Associated with Weight Loss - Non-Clinical](#)
156 [Recommendations](#),”¹⁷ which, once finalized, will provide recommendations for the non-clinical
157 testing to support marketing submissions for these devices.

159 **IV. Clinical Performance Testing Considerations**

160 Generally, non-clinical evaluation does not fully characterize all clinical experience, outcomes,
161 and risks for these devices. We recommend submitters conduct *in vivo* (i.e., clinical) studies to
162 evaluate device safety and effectiveness for new or significantly modified devices with
163 indications associated with weight loss. For novel device designs, feasibility clinical studies can
164 provide important safety and some effectiveness data that can be used to support a pivotal study.
165 Pivotal studies can provide important safety and effectiveness data used to support marketing
166 authorization.

167
168 Devices within the scope of this guidance document are generally considered significant risk
169 devices and subject to all requirements of the Investigational Device Exemptions (IDE)
170 regulation, 21 CFR part 812, for studies conducted in the United States (U.S.). See the FDA
171 guidance titled, “[Significant Risk and Nonsignificant Risk Medical Device Studies](#).”¹⁸ In
172 addition to the requirements of 21 CFR part 812, sponsors of such trials of a device conducted in
173 the U.S. generally must comply with the regulations governing institutional review boards (21
174 CFR part 56) and informed consent (21 CFR part 50).

¹⁶ See reclassification order, available at https://www.accessdata.fda.gov/cdrh_docs/pdf21/DEN210045.pdf.

¹⁷ When final, this guidance will represent FDA’s current thinking on non-clinical testing for medical devices with indications associated with weight loss. Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/medical-devices-indications-associated-weight-loss-non-clinical-recommendations>.

¹⁸ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/significant-risk-and-nonsignificant-risk-medical-device-studies>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

175
176 Obesity represents a heterogeneous disease impacted by demographic, clinical and behavioral
177 factors.¹⁹ Additionally, culture and public health policy can impact weight loss.²⁰ Thus, FDA has
178 encountered challenges about the applicability of foreign effectiveness data to the U.S.
179 population for devices with indications associated with weight loss. Therefore, we recommend
180 that pivotal studies be conducted in the U.S. If foreign data is collected, we recommend that no
181 more than 50% of the pivotal study data be collected from outside the United States (O.U.S.).
182 We also recommend that no more than 20% of the total enrollment population be from one site
183 to avoid the study outcome being dominated by sites with large enrollment.
184

185 When data from clinical investigations conducted O.U.S. are submitted to FDA, the requirements
186 of 21 CFR 812.28 may apply.²¹ 21 CFR 812.28 outlines the conditions for FDA acceptance of
187 clinical data from investigations conducted O.U.S. when submitted to support a premarket
188 submission. For more information, see the FDA guidance “[Acceptance of Clinical Data to
189 Support Medical Device Applications and Submissions: Frequently Asked Questions.](#)”²²
190

191 **A. Study Design**

192 We recommend that pivotal studies to support a weight loss indication be double-blinded,
193 randomized, controlled trials (RCTs). We recommend that additional study staff remain blinded
194 throughout the study (e.g., dieticians, personnel collecting study data).
195

196 We recommend a sham-controlled study as a placebo effect is anticipated. A sham control in a
197 clinical study can provide an important comparator from which to determine the effectiveness of
198 device therapy in comparison to the placebo effect. Therefore, a sham control is beneficial to
199 reduce the uncertainty regarding the treatment effects of the device. We recommend the sham
200 device and/or sham procedure be designed in a way to minimize the subject’s ability to
201 determine whether they have the study device or the sham device. We recommend that
202 submitters consider how blinding will be assessed if using a sham control. We appreciate that a
203 sham control may not be appropriate in all circumstances. If a sham device or sham procedure is
204 not appropriate for a clinical trial design, we recommend a concurrent control arm where the
205 control and treatment groups follow the same lifestyle programs. For all study designs, we
206 recommend standardized dietary and behavioral study aspects between study arms and among
207 centers involved in the study, and that these study aspects be representative of real-world diet
208 and behavior regimens.
209

¹⁹ Jimenez, M. P., Green, M. A., Subramanian, S. V., & Razak, F. (2018). A demographic, clinical, and behavioral typology of obesity in the United States: an analysis of National Health and Nutrition Examination Survey 2011-2012. *Annals of epidemiology*, 28(3), 175–181.e4.

²⁰ Waxman A. WHO Global Strategy on Diet, Physical Activity and Health. *Food and Nutrition Bulletin*. 2004;25(3):292-302.

²¹ This applies to data from clinical investigations that began on or after February 21, 2019 and are submitted to support a premarket submission, including IDEs, PMAs, and 510(k)s.

²² <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/acceptance-clinical-data-support-medical-device-applications-and-submissions-frequently-asked>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

210 For additional information on principles for the design of premarket clinical studies, refer to
211 FDA’s guidance “[Design Considerations for Pivotal Clinical Investigations for Medical](#)
212 [Devices](#).”²³
213

214 **B. Study Duration and Follow-up Schedule**

215 The study should be designed to include adequate follow up to support the indications for use.
216 The follow-up period should also account for the risk posed by device use.

217
218 To support device effectiveness, study duration and the follow-up schedule should be selected
219 with the proposed indication in mind.

- 220 • For a proposed indication of “weight loss,” the duration of device use and primary
221 endpoint should typically demonstrate weight loss at 12 months or more.
- 222 • A proposed indication of “short term weight loss” can typically be supported with a
223 duration of device use and primary endpoint demonstrating weight loss at six months
224 or more, but less than 12 months.
- 225 • Weight loss measured at, or a device that is used for, less than six months could be
226 supportive of a proposed “weight management” indication.
- 227 • Additional follow-up may also be warranted to understand the durability of weight
228 loss. Sometimes a supplemental marketing submission is submitted after these
229 additional follow-up data are collected, for example, to update labeling.
230 Consequently, we recommend consenting patients long enough for any anticipated
231 additional follow-up which may be necessary to support such labeling (or other)
232 modifications.

233
234 To support device safety, study duration and follow-up should be adequate to collect sufficient
235 adverse event information depending on the device design and how it is used. The duration of
236 follow-up needed to support device safety may be longer than that to support effectiveness if
237 warranted due to the risk that the device may pose to patients.
238

239 **C. Inclusion/Exclusion Criteria**

240 As body mass index (BMI) increases, risk of weight-related morbidity and mortality increases.
241 The BMI range for inclusion in a clinical study should be the result of a risk-based decision to
242 ensure that study patients will have an appropriate level of anticipated benefit to offset the risks
243 associated with the device.
244

245 In general, clinical trials of implanted or surgically-placed devices should enroll individuals with
246 a BMI greater than or equal to 35 kg/m², or greater than or equal to 30 kg/m² if accompanied by

²³ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/design-considerations-pivotal-clinical-investigations-medical-devices>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

247 weight-related comorbidities (e.g., type 2 diabetes mellitus (T2DM)).²⁴ In studies of lower risk
248 devices, patients with a BMI of 27 kg/m² with weight-related comorbidities may be included.
249 Higher-risk device studies may warrant additional specification of the BMI range and/or weight-
250 related comorbidities, to ensure that the anticipated benefit outweighs the probable risks.

251
252 Given the risks associated with implanted or surgically-placed devices, patients in studies of such
253 devices should have failed more conservative, first-line weight loss methods such as diet,
254 exercise, and behavior modification.

255
256 Treatment with these medical devices in a clinical study may not be appropriate for certain
257 patients. We recommend that submitters consider the following for the exclusion criteria as
258 applicable:

- 259 • Patients who are unable or unwilling to follow the dietary restrictions specified by the
260 clinical protocol;
- 261 • Altered anatomy (e.g., sleeve gastrectomy);
- 262 • History of dysmotility or delayed gastric emptying;
- 263 • Pregnancy or breastfeeding;
- 264 • Current smokers, because of the contribution of smoking to obesity-linked
265 comorbidities and increased risk of complications;
- 266 • Persons with a history of eating disorder(s), or a serious or uncontrolled psychiatric
267 illness that could compromise understanding or compliance with visits and device
268 maintenance/removal;
- 269 • Active substance abuse;
- 270 • Untreated endocrine or metabolic cause for obesity;
- 271 • Previous gastrointestinal surgery (e.g., bowel resection); and
- 272 • Older patients for whom the risks of the procedure are not acceptable and/or the
273 anticipated lifespan conflicts with the expected period of benefit.

274

D. Patient Demographics

275
276 We recommend that submitters include in their study a representative sample of patients from
277 various demographic groups (e.g., sex, gender, age, ethnic, and racial) in which the prevalence of
278 obesity is highest. FDA recommends that clinical studies for these devices enroll participants that
279 reflect the demographics for clinically relevant populations.

280

²⁴ This recommendation is consistent with the 2018 position statement of the American Society of Metabolic and Bariatric Surgery (ASMBS): Aminian, A., Chang, J., Brethauer, S. A., Kim, J. J. (2018). ASMBS updated position statement on bariatric surgery in class I obesity (BMI 30–35 kg/m²), *Surgery for Obesity and Related Diseases*, 14(8), 1071-1087.

Contains Nonbinding Recommendations

Draft – Not for Implementation

281 For more information regarding the evaluation and reporting of age, race, ethnicity and sex-
282 specific data in medical device clinical studies, see FDA’s guidances “[Evaluation of Sex-
283 Specific Data in Medical Device Clinical Studies](#)”²⁵ and “[Evaluation and Reporting of Age-,
284 Race-, and Ethnicity-Specific Data in Medical Device Clinical Studies](#).”²⁶
285

286 **E. Treatment Parameters/Protocol**

287 The study-specific treatment protocol should minimize risk to patients. The protocol should not
288 only consider the risks associated with the device and device placement, but any additional risk
289 that may be applicable to all patient populations included in the study. For example, if submitters
290 choose to include patients with certain comorbidities (e.g., T2DM), the protocol should explain
291 how these patients will be protected from complications that may arise due to their disease.
292

293 Specifically, when designing trials that include patients with T2DM, we recommend that a safety
294 monitoring plan be included in the protocol to detect and manage hypoglycemia or continued
295 uncontrolled hyperglycemia. The management plan should consider an algorithm for the
296 lowering or elimination of oral hypoglycemics or insulin based on fasting glucose levels and/or
297 glycated hemoglobin (HbA1c)²⁷ (for patients who lose clinically significant amounts of weight).
298

299 For a device with novel technology and/or with an undefined risk profile, it may also be
300 appropriate to prospectively define stopping rules in the study protocol and/or initially enroll a
301 limited number of patients in a phased manner to better manage risk.
302

303 If the device is a permanent implant, the study design should include considerations for how a
304 device should be explanted if warranted or requested during or at termination of the study.
305 Considerations should include, at a minimum, removal instructions and a plan for tracking
306 reasons for device explant, including association with any adverse events as noted in Section
307 IV.F below. There should also be evidence that removal instructions in device labeling are
308 sufficient to safely remove the device if explant is warranted. Removal instructions should be
309 evaluated during the course of the clinical study if devices are explanted from patients.
310

311 Throughout the study, participants should receive the standard of care, including medication and
312 monitoring for comorbidities such as hypertension, dyslipidemia, and glycemic control.
313

314 **F. Safety Endpoints and Data**

315 The primary safety endpoint should be reporting of all device- and procedure-related adverse
316 events, as FDA intends to consider all adverse events in our assessment of the premarket

²⁵ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/evaluation-sex-specific-data-medical-device-clinical-studies-guidance-industry-and-food-and-drug>.

²⁶ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/evaluation-and-reporting-age-race-and-ethnicity-specific-data-medical-device-clinical-studies>.

²⁷ HbA1c (glycated hemoglobin) is a term commonly used in relation to diabetes - the higher the HbA1c, the greater the risk of developing diabetes-related complications.

Contains Nonbinding Recommendations

Draft – Not for Implementation

317 submission. Additional safety assessments may be warranted based on the design and principles
318 of operation of the specific device.
319

320 **G. Effectiveness Endpoints and Data**

321 Demonstrated weight loss should be based on percent total body weight loss (% TBWL),²⁸ which
322 is typically captured in a clinical study with co-primary effectiveness endpoints that include:

- 323 • a hypothesis with a pre-specified superiority margin of the mean % TBWL over
324 control; and
- 325 • a performance goal for a responder rate based on individual subject success.

326
327 FDA recommends a pre-specified superiority margin for mean % TBWL be included in the
328 clinical protocol depending on the indication being sought in the premarket submission:

- 329 • For an indication of “weight loss,” we recommend at least a 5% superiority margin of
330 the mean % TBWL over the control. However, the minimum value over the control
331 arm should be appropriate for the risk associated with device use and any device-
332 related procedures.
- 333 • For an indication of “limited weight loss,” we recommend at least a 2% superiority
334 margin of the mean % TBWL over the control. However, the minimum value over the
335 control arm should be appropriate for the risk associated with device use and any
336 device-related procedures.
- 337 • For an indication of “weight management,” a superiority margin of less than 2% may
338 be supportive if additional benefit is measured (i.e., responder rate endpoint is met).
339 However, the benefit should be appropriate for the risk associated with device use and
340 any device-related procedures.

341
342 For the responder rate, we recommend that at least 50% of treated patients achieve at least 5%
343 TBWL for any indication associated with weight loss (i.e., weight loss, limited weight loss,
344 weight reduction, weight management, or obesity treatment).
345

346 For an indication of “obesity treatment,” we recommend endpoint(s) demonstrating clinical
347 benefits in addition to weight loss alone. Support for additional benefits should be appropriately
348 powered in the study design.
349

350 We recommend submitters consider the following secondary effectiveness endpoints:

- 351 • Percent excess weight loss (% EWL);²⁹
- 352 • Change in weight;

²⁸ For the purposes of this guidance, FDA defines % TBWL = [(initial weight – final weight)/initial weight] × 100%.

²⁹ For the purposes of this guidance, FDA defines % EWL = [(initial weight – weight to be at a BMI of 25)/initial weight] × 100%.

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 353 • Change in BMI; and
354 • Change in waist circumference.

355
356 We also recommend that submitters consider including patient-reported outcomes (PROs)³⁰ and
357 patient preference information (PPI)³¹. The value patients associate with the treatment, their
358 willingness to accept the risk of this treatment to achieve the benefit, the treatment’s ability to
359 improve the patient’s overall quality of life, and the patient’s ability to understand the benefits
360 and risks of the treatments are important factors in evaluating device benefit.

361
362 Changes in common weight-related comorbidities are often secondary endpoints in studies of
363 devices with indications associated with weight loss. If any of the secondary endpoint analyses
364 are intended to support the indications for use or to describe device performance in the labeling
365 (e.g., comparing treatment and control groups using p-values or confidence intervals), we
366 recommend pre-specifying this intention in the study protocol and providing a detailed
367 description of the statistical methods planned to follow. The study should be powered
368 appropriately to evaluate such changes, if comparative statements are intended to be made in the
369 labeling.

370

371 **H. Adverse Events**

372 We recommend that all adverse event data be collected during the study and that events be
373 adjudicated as to whether they are device- and/or procedure- related. In general, we recommend
374 that studies have a data safety monitoring board (DSMB) and establish an endpoint
375 assessment/adjudication committee. We refer the submitter to the FDA guidance “[Establishment
376 and Operation of Clinical Trial Data Monitoring Committees](#)”³² for more information.
377 Independent data monitoring committees help to ensure the safety of enrolled participants as
378 follows:

- 379 • The committee can provide a comparative assessment of accumulating safety and
380 effectiveness data to inform recommendations to the study sponsor whether to
381 continue, modify, or stop the study;
- 382 • Potential complications may warrant robust study oversight from a third party that is
383 advisory to the study sponsor; and

³⁰ See FDA guidances “Principles for Selecting, Developing, Modifying, and Adapting Patient-Reported Outcome Instruments for Use in Medical Device Evaluation” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/principles-selecting-developing-modifying-and-adapting-patient-reported-outcome-instruments-use>, and “Patient-Reported Outcome Measures: Use in Medical Product Development to Support Labeling Claims,” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/patient-reported-outcome-measures-use-medical-product-development-support-labeling-claims>.

³¹ See FDA guidance “Patient Preference Information - Voluntary Submission, Review in Premarket Approval Applications, Humanitarian Device Exemption Applications, and De Novo Requests, and Inclusion in Decision Summaries and Device Labeling,” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/patient-preference-information-voluntary-submission-review-premarket-approval-applications>.

³² <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/establishment-and-operation-clinical-trial-data-monitoring-committees>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 384 • Unbiased adjudication of adverse events reduces the uncertainty in study safety
385 outcome data.

386
387 We recommend an adverse event classification modeled after the Clavien-Dindo Classification
388 of Surgical Complications,³³ shown in Table 2, where the severity of each adverse event is
389 graded based on the treatment used to address the event.

390
391 **Table 2. Adverse event classification for clinical studies**

Grade	Definition
Grade I	Any deviation from the normal treatment course without the need for surgical, endoscopic, and radiological interventions. Includes all over-the-counter pharmacological interventions and non-narcotic prescription pain medications (including anti-emetics, antipyretics, analgesics, diuretics, electrolytes, physiotherapy, and bedside wound care)
Grade II	Requiring pharmacological treatment with prescription drugs (excluding non-narcotic pain medications in Grade I), the administration of intravenous fluids, blood transfusions, or total parenteral nutrition
Grade III	Requiring surgical, endoscopic, or radiological interventions
Grade IV	Life-threatening complications requiring intensive care/intensive care unit management (including single and multiorgan dysfunction, and central nervous system complications)
Grade V	Death

392
393 The classification scheme identified in Table 2 focuses on deviations from the normal treatment
394 course for a device. For example, the normal treatment course for a device may include use of
395 concomitant medications, and additional therapy (e.g., anti-emetics, pain medication) typically
396 provided as part of the practicing physician’s treatment plan. While concomitant medications are
397 not considered as adverse events per this classification scheme, FDA does consider such as part
398 of the overall benefit-risk determination for a device, as described in Table 4 in Section V.A.

399
400 A single type of adverse event can be categorized into different grades, depending on the
401 treatment required for resolution. For example, vomiting can be resolved with over-the-counter
402 medication (Grade I), or vomiting can require administration of intravenous fluids (Grade II).
403 The grades are to be considered mutually exclusive, and together the grades should cover all
404 event outcomes. All events that fit into a single grade are of approximately equal severity/risk to
405 the patient.

406

³³ Dindo, D., Demartines, N., & Clavien, P. A. (2004). Classification of surgical complications: a new proposal with evaluation in a cohort of 6336 patients and results of a survey. *Annals of surgery*, 240(2), 205.

Contains Nonbinding Recommendations

Draft – Not for Implementation

407 We recommend submitters present adverse event information to FDA in their premarket
408 submission as follows:³⁴

- 409 • Tabulate all adverse events and categorize as device-related, procedure-related, or not
410 related to the device or procedure and categorize all adverse events as explained in
411 Table 2;
- 412 • Tabulate all serious adverse events (SAEs) and categorize as device-related,
413 procedure-related, or not related to the device or procedure and categorize all SAEs as
414 explained in Table 2;
- 415 • Identify any and all unanticipated adverse device effects;
- 416 • Provide the time to onset as well as duration for all gastrointestinal-associated device-
417 and/or procedure-related adverse events, including resolution status; and
- 418 • Tabulate all unanticipated device removals and the reason for removal.

419
420 We recommend the use of PRO instruments to assess non-serious adverse events using validated
421 tools such as the gastrointestinal symptom scales included in the National Institutes of Health
422 (NIH) PRO Measurement Information System (PROMIS).³⁵
423

424 **I. Statistical Analysis Considerations**

425 **(1) Sample Size**

426 For pivotal studies, we recommend that co-primary effectiveness endpoints include a hypothesis
427 with a pre-specified superiority margin for percent total body weight loss and a performance goal
428 for a responder rate. The number of patients should be the maximum of sample sizes calculated
429 based on the co-primary endpoints considering anticipated loss to follow-up; however, additional
430 patients should be enrolled to assess device safety to support premarket submission. In general,
431 calculations should be based on two-sided tests of significance at the 5% level and at least 80%
432 power. Effect sizes for the calculations should represent clinically meaningful differences.
433

434 **(2) Analysis Methods**

435 Endpoints should be analyzed based on the intent-to-treat (ITT) population, defined as patients
436 that were enrolled and randomized into the study, regardless of whether the patients received the
437 treatment to which they were randomized.
438

³⁴ As described in Section III.B.(4), of FDA’s draft guidance, “Medical Devices with Indications Associated with Weight Loss - Non-Clinical Recommendations,” FDA recommends that the adverse event information in this list also be included in the device’s labeling. Available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/medical-devices-indications-associated-weight-loss-non-clinical-recommendations>.

³⁵ Spiegel, B. M., Hays, R. D., Bolus, R., Melmed, G. Y., Chang, L., Whitman, C., ... & Khanna, D. (2014). Development of the NIH patient-reported outcomes measurement information system (PROMIS) gastrointestinal symptom scales. *The American journal of gastroenterology*, 109(11), 1804.

Contains Nonbinding Recommendations

Draft – Not for Implementation

439 The analysis of % TBWL should use analysis of variance (ANOVA) or analysis of covariance
440 (ANCOVA) with baseline weight as a covariate in the model.

441
442 Response rates should be compared between the treatment and control groups using statistical
443 methods appropriate for categorical data. A sensitivity analysis should be conducted that
444 considers patients who are treated, drop out, and do not have complete post-baseline data as
445 treatment failures. Additionally, a tipping point analysis for binary response variables should be
446 considered.

447
448 Type I error should be controlled across all clinically relevant secondary effectiveness endpoints
449 intended for product labeling.

450

451 **(3) Missing Data**

452 **a. Efforts to reduce missing data**

453 We recommend you describe the efforts that will be used during the course of the study to
454 monitor and minimize the incidence of patient dropouts, such as monitoring activities, special
455 incentives to patients for study compliance, methods to remind patients of scheduled visits, and
456 specific efforts to contact patients who miss their visit (e.g., telephone calls, postcards, contact
457 next-of-kin).

458

459 **b. Document reasons for missing data**

460 We recommend you identify the steps to document:

- 461 • the reason for each missed visit, e.g., complications, difficulty getting transportation to
462 the site; and
- 463 • the reason for each dropout, e.g., seeking alternate therapy, complications or intolerance
464 to the device, dissatisfaction with the device, moved away.

465

466 To permit a complete and detailed accounting of all study patients, we recommend you collect
467 complete information during the study because loss to follow-up jeopardizes the conclusions that
468 can be made about the long-term safety and effectiveness of a device.

469

470 **c. Handling missing primary endpoint data**

471 To allow for a true ITT analysis, we recommend obtaining body weight measurements in all
472 patients who prematurely withdraw from studies near the calendar date at which they were
473 scheduled. This will reduce uncertainty in the ultimate outcome of the study by having a data
474 measurement at the primary effectiveness endpoint rather than imputing the measurement. For
475 example, a patient who withdraws from a 12-month study after six months of treatment should
476 have a body weight measurement at the time he or she would have completed 12 months of study
477 participation. If this is not possible, we recommend conducting sensitivity analyses to determine
478 the best mechanism to account for missing data.

479

Contains Nonbinding Recommendations

Draft – Not for Implementation

480 **d. Sensitivity analyses**

481 Sensitivity analyses employing imputation strategies should assess the effect of dropouts on the
482 results. The imputation strategy should be prespecified and should consider the expected dropout
483 patterns and the time-course of weight changes in the treatment group. No imputation strategy
484 will work for all situations, particularly when the dropout rate is high, so a primary study
485 objective should be to keep missing values to a minimum. We recommend multiple imputation
486 when a “missing at random” assumption is plausible. For early exit due to adverse events or
487 ineffectiveness of the device, we recommend you use “unfavorable clinical outcome” to impute
488 missing data.
489

490 **(4) Subgroup Analyses**

491 We recommend submitters conduct gender and sex-based subgroup analyses. We recommend
492 submitters conduct subgroup analyses based on race and ethnicity as the prevalence of obesity
493 varies among these groups in the U.S. population.³⁶ If the study includes sites O.U.S. then we
494 recommend conducting a U.S. subgroup analysis.
495

496 **J. Pediatric Studies**

497 Planning clinical trials for pediatric patients includes additional considerations beyond those of
498 adult patients, such as ethical issues of studying a more vulnerable patient population and an
499 altered benefit-risk profile because of potential interference of a medical device with physical
500 growth and maturation. Consistent with the FDA guidance “[Premarket Assessment of Pediatric
501 Medical Devices](#),”³⁷ FDA considers patients below 22 years of age to be pediatric (that is, from
502 birth up to but not including the 22nd birthday) for medical device studies.
503

504 The increased prevalence of children being overweight or having obesity, emphasizes an unmet
505 need to provide therapy to children who have a disease that impacts their health, quality of life,
506 and psychosocial factors. FDA remains open to considering risk-based clinical study designs and
507 intends to consider both the benefits and risks to adolescent study participants when determining
508 the amount of benefit-risk evidence needed before initiation of an adolescent weight-loss device
509 study.
510

511 We recommend using the U.S. Centers for Disease Control and Prevention (CDC) National
512 Center for Health Statistics definitions for classifying pediatric-aged patients as overweight or

³⁶ <https://www.niddk.nih.gov/health-information/health-statistics/overweight-obesity>.

³⁷ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/premarket-assessment-pediatric-medical-devices>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

513 obese and the American Heart Association recommendation for severe obesity based on age- and
514 sex-matched BMI cutoffs as follows:^{38, 39, 40, 41}

- 515 • BMI-for-age between the 85th and 95th percentile is overweight;
- 516 • BMI-for-age at or above the 95th percentile is obesity; and
- 517 • BMI \geq 120% of the 95th percentile or an absolute BMI \geq 35 kg/m², whichever is lower
518 based on age and sex is severe obesity.

519
520 FDA developed the following recommendations considering outcomes from the 2005 FDA
521 Pediatric Advisory Committee (PAC) meeting on weight loss device clinical trial designs for
522 pediatric patients,⁴² changes in the field of childhood obesity since the PAC's
523 recommendations,⁴³ and input from external experts, including clinicians. Additionally, the
524 following recommendations are intended to supplement and not supersede those discussed in the
525 FDA guidance "[Premarket Assessment of Pediatric Medical Devices](#)."⁴⁴ These recommendations
526 are in addition to those discussed elsewhere in this document for adult patients.

527
528 Recommendations specific for pediatric patients include:

- 529 1. In general, the device should not be studied in the pediatric population until enough data
530 has been obtained to show that the study does not involve greater than minimal risk.^{45,46}
531 Additionally, if the device is a permanent implant, sufficient data should exist to support
532 anticipated benefit in the pediatric population.⁴⁷ Other sources of data, including animal

³⁸ Kelly, A. S., Barlow, S. E., Rao, G., Inge, T. H., Hayman, L. L., Steinberger, J., ... & Daniels, S. R. (2013). Severe obesity in children and adolescents: identification, associated health risks, and treatment approaches: a scientific statement from the American Heart Association. *Circulation*, 128(15), 1689-1712.

³⁹ Gulati, A. K., Kaplan, D. W., & Daniels, S. R. (2012). Clinical tracking of severely obese children: a new growth chart. *Pediatrics*, 130(6), 1136-1140.

⁴⁰ Ogden, C. L., & Flegal, K. M. (2010). Changes in terminology for childhood overweight and obesity. *Age*, 12(12).

⁴¹ Flegal, K. M., Wei, R., Ogden, C. L., Freedman, D. S., Johnson, C. L., & Curtin, L. R. (2009). Characterizing extreme values of body mass index-for-age by using the 2000 Centers for Disease Control and Prevention growth charts. *The American journal of clinical nutrition*, 90(5), 1314-1320.

⁴² FDA Pediatric Advisory Committee, Development of Trials to Assess the Safety and Efficacy Relevant to Scientific and Ethical Issues Surrounding Trials for Pediatric Devices for Weight Loss. Gaithersburg, MD. Meeting materials can be accessed at <https://wayback.archive-it.org/7993/20170403222257/https://www.fda.gov/ohrms/dockets/ac/oc05.html#Pediatric>.

⁴³ Marrone A.K., Venkataraman-Rao P., Gottschalk L. (2021). Food and Drug Administration insights on clinical study of weight-loss devices intended for adolescent patients. *Pediatric Obesity*, e12768.

⁴⁴ <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/premarket-assessment-pediatric-medical-devices>.

⁴⁵ In general, the 2005 FDA PAC recommended that "devices, especially implants, should not be studied in the pediatric population until enough data has been gained from adult study and use." For more information, see https://wayback.archive-it.org/7993/20170404062450/https://www.fda.gov/ohrms/dockets/ac/05/minutes/2005-4179m_summary.pdf.

⁴⁶ In general, the 2005 FDA PAC recommended that "A staged introduction should be used when studying devices for obesity in the pediatric population. Namely, after adequate information is available in adult populations, the device can be studied in the older adolescent group (12 or 14 to 17). Sufficient experience and data should be collected before studying the device in patients younger than this." *Ibid*.

⁴⁷ In general, the 2005 FDA PAC recommended that "post-approval data should be collected through 5 years" and "parties should be encouraged to have registries for long-term implants, which follow patients for 5-10 years." *Ibid*.

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 533 or other relevant modeling and simulation data, may preclude or mitigate the need to
534 preliminarily collect data on older populations. This may be especially relevant when
535 designing clinical investigations to meet the more immediate needs of patients, such as
536 younger adolescents, experiencing co-morbidities associated with the severe end of the
537 obesity spectrum.
- 538 2. If the device is a permanent implant, risk associated with potential explantation of the
539 permanent implant should be well defined.
 - 540 3. Pediatric patients should have a documented history of failing to achieve weight-loss
541 goals with lifestyle modification before enrollment into a clinical study for these devices.
542 In general, patients should have participated in a comprehensive, multi-disciplinary
543 pediatric weight management program for at least six months without adequate results.
 - 544 4. Studies should have a lead-in period that allows for adequate time for the clinical team to
545 get to know the patient, for the failure of adequate therapy programs to be documented,
546 for the patient to understand the therapy and its impact, and for the patient’s ability to
547 comply with diet, protocol, and other considerations (e.g., psychosocial comorbidities) to
548 be assessed.⁴⁸
 - 549 5. FDA considers the risk profile of the device for the appropriate study population in a
550 pediatric clinical study. Table 3 illustrates recommended percentiles for BMI-for-age for
551 inclusion of adolescent patients into a study for a device with indications associated with
552 weight loss. Generally, higher risk devices should have the potential for greater benefit,
553 as indicated by the percentiles for BMI-for-age in Table 3. If the submitter believes that
554 the device is low-risk, FDA encourages discussion of a risk-based justification for
555 inclusion of study patients with lower BMI-for-age percentiles.

556
557 **Table 3.** Recommended percentiles (%ile) for BMI-for-age values for inclusion of adolescent
558 patients into a study for a device with an indication associated with weight loss. Risk-dependent
559 value should fall within specified ranges.

	Comorbidity	No comorbidity
Temporary ⁴⁹ device	≥85 th %ile BMI	≥95 th %ile BMI
Permanent ⁵⁰ device	≥85 th %ile to 120% of the 95 th %ile BMI	≥95 th %ile to 140% of the 95 th %ile BMI

560

⁴⁸ In general, the 2005 FDA PAC recommended that “studies should have a lead-in period during which the physician team got to know the patient and it could be documented that the patient had failed adequate conservative therapy programs and to ensure the patient’s ability to comply with diet, protocol, etc.” *Ibid.*

⁴⁹ For the purposes of interpreting this table, a temporary device is intended to be implanted or used for a pre-determined, limited amount of time (for example: a six-month intragastric balloon). A permanent device is one that is implanted without intention to remove or one that permanently alters the patient’s anatomy and/or physiology. For the purposes of device classification procedures, the definition of an implant is provided in 21 CFR 860.3.

⁵⁰ *Ibid.*

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 561 6. A study endpoint of less than 12 months is likely not appropriate to evaluate a permanent
562 device in the pediatric population, as these patients are still growing and maturing.⁵¹
- 563 7. Obesity-related comorbidities that should be considered for inclusion include:⁵²
- 564 • Obstructive sleep apnea;
- 565 • Prediabetes;
- 566 • T2DM;
- 567 • Uncontrolled hypertension;
- 568 • Orthopedic complications;
- 569 • Pseudotumor cerebri;
- 570 • Non-alcoholic steatohepatitis (NASH);
- 571 • Polycystic ovary syndrome (PCOS); and
- 572 • Hyperlipidemia/dyslipidemia.
- 573 8. Exclusion criteria should include:
- 574 • Uncontrolled psychiatric conditions;
- 575 • Patients that are ill-equipped or unwilling to change behavior;
- 576 • Patients who are unwilling to undergo the intervention themselves;
- 577 • Patients with anatomical issues that may put them at unreasonable risk;
- 578 • Patients with connective tissue disorders that may result in tissue breakdown, if the
579 device is an implant or changes anatomy; and
- 580 • Developmentally disabled patients who cannot follow recommendations.
- 581 9. To determine suitability for participating in a clinical study, maturity level and
582 psychosocial comorbidities should be assessed by a specialist trained in psychology and
583 in discussing mental health issues, stigma, bias, bullying, binge-purge behaviors,
584 readiness for change, and other related considerations.
- 585 10. Patients should be screened for known genetic causes of obesity such as Prader-Willi
586 Syndrome.⁵³ For these patients, as well as those with hypothalamic obesity related to

⁵¹ In general, the 2005 FDA PAC recommended that “Premarket data should be collected for 2 years although patients should be consented/assented for 5 years.” For more information, see https://wayback.archive-it.org/7993/20170404062450/https://www.fda.gov/ohrms/dockets/ac/05/minutes/2005-4179m_summary.pdf.

⁵² Obesity-related comorbidities listed are also applicable to adult study populations. These comorbidities are listed in this section of this guidance document due to the relevant general recommendations from the 2005 FDA PAC: 1) Long-term implant devices should be studied in patients with significant disease, i.e., those who are in the 99th percentile for BMI-for-age, and have at least one significant comorbidity, such as sleep apnea, diabetes, pseudotumor cerebri, or NASH (Non-Alcoholic Steatohepatitis); and 2) Comorbidity reduction or resolution would be an important secondary effectiveness endpoint although the study would need to be powered appropriately to evaluate such changes. *Ibid.*

⁵³ In general, the 2005 FDA PAC recommended that “patients should be screened for known genetic causes of obesity and for Prader Willi, and if included in the study, should be evaluated separately.” *Ibid.*

Contains Nonbinding Recommendations

Draft – Not for Implementation

587 craniopharyngioma surgery that are about eight years old and above, inclusion into a
588 study could be considered, though FDA recommends that the submitter consider
589 separately evaluating the data for this subpopulation.

590 11. As for adult studies, clinically meaningful weight loss may be defined by % TBWL that
591 should be linked to the health risk in the desired pediatric patient population. Consistent
592 with clinical guidelines based on cardiometabolic risk,⁵⁴ we consider at least a 5-10 %
593 TBWL clinically meaningful, and these values could be applicable to the pediatric
594 population. However, linear growth should be considered when assessing changes in
595 body weight of children and adolescents. Thus, the primary effectiveness parameter could
596 be a function of the change in %BMI-for-age and/or % TBWL. This should depend on
597 what is most clinically meaningful in the desired patient population considering age, BMI
598 range, and any additional disease factors (e.g., associated comorbidities). Additionally,
599 endpoint(s) should be able to demonstrate a positive outcome on the disease status (e.g.,
600 change in class of obesity).

601 12. If applicable, comorbidity reduction or resolution should be a secondary effectiveness
602 endpoint.

603 13. The overall clinical study duration and follow-up should be justified considering the
604 anticipated benefit and device risk. However, for devices that result in the modification of
605 anatomy or involve a permanent implant, we recommend that premarket evaluation
606 include follow-up for two years to account for weight loss durability. Patients should be
607 consented or assented, as applicable, for five years to allow for longer-term follow-up
608 post-marketing. Parental permission should be obtained when applicable.

609 14. For a device that is temporary, durability of device-effect should be measured at least six
610 months post device use unless a shorter assessment period is justified.

611 15. Height measurements should be obtained from a wall-mounted stadiometer by study
612 personnel trained in its use. A bone age study to obtain radiographic imaging of the
613 growth plates can also be considered.

614
615 Other clinically relevant issues to consider when designing a pediatric study include
616 endocrinologic causes of obesity, assessing neuropsychiatric symptoms and/or psychosocial
617 environment, compliance, nutritional issues, and reproduction issues. We recommend addressing
618 and/or monitoring these issues as appropriate.

619
620 We encourage submitters to utilize FDA's [Q-Submission Program](#) to ensure that the pediatric
621 study protocol addresses safety concerns depending on the facts and circumstances of the device
622 and study.

623

⁵⁴ Jensen, M. D., Ryan, D. H., Apovian, C. M., Ard, J. D., Comuzzie, A. G., Donato, K. A., & Yanovski, M. (2014). AHA/ACC/TOS guideline for the management of overweight and obesity in adults: a report of the American College of Cardiology/American Heart Association Task Force on Practice Guidelines and The Obesity Society. *Circulation*, 129(25 Suppl 2), S102-138.

624 V. Benefit-Risk Considerations

625 A. Benefit-Risk

626 FDA evaluates whether a device has a reasonable assurance of safety and effectiveness during
627 the PMA review, or whether general or general and special controls provide such assurance for a
628 device in a De Novo classification, or whether it is substantially equivalent to a valid predicate in
629 510(k) review, by weighing any probable benefit to health from the use of the device against any
630 probable risk of injury or illness from such use,⁵⁵ or assessing the benefit-risk profile of a device
631 as compared to a valid predicate,⁵⁶ among other relevant factors. To aid in this process,
632 submitters include valid scientific evidence, including one or more clinical investigations, where
633 appropriate, and/or non-clinical information, which FDA reviews to determine, among other
634 things, whether the device will have the effect it purports or is represented to have under the
635 conditions of use prescribed, recommended, or suggested in the labeling of the device.⁵⁷

636
637 When assessing the benefits of devices, FDA considers the types of benefits, the magnitude of
638 benefits, the probability of patients experiencing one or more benefits, and the duration of
639 effects.⁵⁸ When assessing the risks of devices, FDA considers severity, type, number, and rate of
640 harmful events associated with use of the device or procedure associated with the device,
641 probability of harmful events, and duration of harmful events. Additional factors considered
642 when assessing the probable benefits and risks of devices include uncertainty⁵⁹ surrounding the
643 benefit and risk, patient-centric assessments and PROs, characterization of the disease or
644 condition, patient preferences,⁶⁰ availability of alternate treatments, risk mitigation, device-type
645 post-market data, and novel technology for addressing unmet medical needs.

646
647 Specific to devices with indications associated with weight loss, important considerations include
648 the factors listed in Table 4.
649

⁵⁵ The criteria for determining the safety and effectiveness of a device are set forth in section 513(a) of the Federal Food, Drug, and Cosmetic Act (FD&C Act) and 21 CFR 860.7.

⁵⁶ See FDA guidance “Benefit-Risk Factors to Consider When Determining Substantial Equivalence in Premarket Notifications (510(k)) with Different Technological Characteristics,” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/benefit-risk-factors-consider-when-determining-substantial-equivalence-premarket-notifications-510k>.

⁵⁷ Section 513(a)(3)(A) of the FD&C Act.

⁵⁸ See FDA guidance “Factors to Consider When Making Benefit-Risk Determinations in Medical Device Premarket Approval and De Novo Classifications,” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/factors-consider-when-making-benefit-risk-determinations-medical-device-premarket-approval-and-de>.

⁵⁹ See FDA guidance “Consideration of Uncertainty in Making Benefit-Risk Determinations in Medical Device Premarket Approvals, De Novo Classifications, and Humanitarian Device Exemptions,” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/consideration-uncertainty-making-benefit-risk-determinations-medical-device-premarket-approvals-de>.

⁶⁰ See FDA guidance “Patient Preference Information - Voluntary Submission, Review in Premarket Approval Applications, Humanitarian Device Exemption Applications, and De Novo Requests, and Inclusion in Decision Summaries and Device Labeling,” available at <https://www.fda.gov/regulatory-information/search-fda-guidance-documents/patient-preference-information-voluntary-submission-review-premarket-approval-applications>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

650
651 **Table 4.** Factors considered as part of the benefit-risk evaluation for devices with indications
652 associated with weight loss

Factor	Example(s)
Assessment of Benefits from a Clinical Study	
Weight loss	amount of weight loss attributed to the device, proportion of patients experiencing weight loss, and durability of weight loss
Changes in comorbidities	improvements in cardiometabolic risk factors, as well as other obesity-related comorbidities (e.g., clinically significant reduction in HbA1c, hypertension, and/or dyslipidemia), reduction in medication(s)
Other benefit	improvement in quality of life
Assessment of Risks from a Clinical Study	
Device- and procedure-related adverse events	seriousness, severity, types, numbers, rates, duration, resolution of adverse events and exacerbation of pre-existing conditions
Effects of the device	permanent implantation, anatomic changes, restriction of future treatment options, reversibility limitations, effect on drug and/or nutrient absorption
Clinical treatments/procedures related to the device	risk associated with expected concomitant medications or therapies, rate of early device removal due to patient request, risks related to placement/removal procedures, risks related to procedures necessary to diagnose adverse events, hospitalization (need, duration, and reason for)
Additional Factors	
Evaluation matrices decision aid⁶¹	extent of weight loss and duration of device use versus prevalence and severity of adverse events reported in a clinical study
Uncertainty	uncertainty resulting from study design, study conduct, potential for sham effect, and range of confidence intervals
Additional clinical data	studies from outside the United States, feasibility studies, real-world evidence, use of the device repeatedly or in sequence
Additional considerations	availability of alternative therapies, risk mitigation measures, patient preferences

653
654 There is a wide range of technology and techniques being attempted for devices with indications
655 associated with weight loss. These different approaches can translate into different impacts or
656 outcomes, such as duration of device implantation, adverse event profiles, and different amounts
657 of weight loss. As innovators conceive and develop the next generation of devices with plans to
658 market such devices in the U.S., the recommendations below explain how FDA intends to

⁶¹ The evaluation matrices are applicable to devices with indications outlined in Table 5.

Contains Nonbinding Recommendations

Draft – Not for Implementation

659 consider, in the context of premarket submission decision, adverse events in light of varying
660 degrees of benefit (specifically extent of weight loss and duration of device use).
661

662 **B. Use of Modified Clavien-Dindo to Assess Risk**

663 As described in Section IV.HH, we recommend an adverse event classification modeled after the
664 Clavien-Dindo Classification of Surgical Complications,⁶² where the severity of each device- and
665 procedure-related adverse event is graded based on the treatment used to address the event (*See*
666 Table 2). The Clavien-Dindo Classification was chosen due to its wide use among physicians as
667 a reliable and reproducible system for reporting surgical complications. Modifications to the
668 Classification system were adapted to make it more relevant for weight loss device-related
669 complications.

670
671 We highlight the differences from the original Clavien-Dindo Classification as well as relevant
672 considerations in the following summation:

- 673 • Grade I was adapted to include over-the-counter medications and non-narcotic
674 prescription pain medications.
- 675 • Grade II includes all other prescription medications and the administration of
676 intravenous fluids.
- 677 • Like the original Clavien-Dindo Classification scheme, length of hospital stay is not
678 included, since practices vary between medical centers and unexpected
679 hospitalization typically occurs in combination with other therapies that are captured
680 by the classification. However, FDA intends to consider seriousness and the need,
681 duration of, and reason for hospitalization when making our overall benefit-risk
682 determination for these devices.
- 683 • Diagnostic procedures, such as diagnostic endoscopies, are not included, because an
684 adverse event discovered by a diagnostic procedure would be classified by the
685 treatment needed for the adverse event. However, FDA intends to consider the risk of
686 diagnostic procedures that may be used to diagnose device- or procedure-related
687 adverse events when making our overall benefit-risk determination for these devices.
- 688 • Regarding Grade II, a patient’s need for blood transfusions and total parenteral
689 nutrition (TPN) would be indicative of more serious adverse events in comparison to
690 prescription medication use; however, the associated adverse events are likely to
691 include additional treatments defined as Grade III or Grade IV, and the grades of
692 those additional treatments would also be captured.
- 693 • Devices can electively be removed prior to the end of their intended course of therapy
694 for reasons other than adverse events included in the Adverse Event Classification
695 described in Table 2. These reasons could be at patient request. These events are not
696 captured in the Adverse Event Classification, but FDA intends to consider early
697 device removal when making our overall benefit-risk determination for these devices.

⁶² Dindo, D., Demartines N., Clavien P.A. (2004) Classification of surgical complications: a new proposal with evaluation in a cohort of 6336 patients and results of a survey. *Ann Surg.* 240:205–213.

698

699

700

C. Balancing Weight Loss and Adverse Events for an Indication of Weight Loss

701 FDA’s assessment of tolerability of adverse events in light of varying degrees of weight loss for
702 devices specifically with a weight loss indication have been developed considering:

- 703 • Outcomes from the 2012 Gastroenterology-Urology Devices Panel on general issues
704 related to obesity treatment devices;⁶³
- 705 • Feedback from external experts, including clinicians; and
- 706 • The public comments submitted to Docket No. FDA-2019-N-4060 in response to a
707 discussion paper outlining concepts discussed below.

708

709 As described in Sections IV.GB and IV.G, indications for weight loss depend on the extent and
710 duration of weight loss demonstrated in a clinical study. For devices used for less than six
711 months, or having less benefit than that outlined in Table 5, a weight management indication
712 may be appropriate. An obesity treatment indication should be supported by clinical benefits in
713 addition to weight loss alone.

714

715 Table 5 summarizes four weight loss indication categories based on the amount of weight loss
716 observed in a clinical study and the duration of device use.

717

718 **Table 5.** Weight loss indication categories

Indication	Demonstrated Weight Loss		Duration of Device Use
	Superiority Margin % TBWL Over Control	Responder Rate % patients achieving ≥5% TBWL	
Short-Term Limited Weight Loss	≥2% and <5%	50%	6 months to <12 months
Limited Weight Loss	≥2% and <5%	50%	≥12 months
Short-Term Weight Loss	≥5%	50%	6 months to <12 months
Weight Loss	≥5%	50%	≥12 months

719

720 For the categories in Table 5, the duration of device use depends on the characteristics of device
721 use. It may depend on the time period over which the device is used and/or the time period over
722 which weight loss is measured, as follows:

⁶³ 2012 Materials of the Gastroenterology-Urology Devices Panel can be accessed at <https://wayback.archive-it.org/7993/20170113191551/http://www.fda.gov/AdvisoryCommittees/CommitteesMeetingMaterials/MedicalDevices/MedicalDevicesAdvisoryCommittee/Gastroenterology-UrologyDevicesPanel/ucm286235.htm>.

Contains Nonbinding Recommendations

Draft – Not for Implementation

- 723
- 724
- 725
- 726
- 727
- 728
- 729
- 730
- 731
- 732
- 733
- 734
- For an implantable device, the duration of device use is the total time that the device is inside the body. For example, for an intragastric balloon that is in the stomach for 6 months and then removed, the duration of device use would be 6 months.
 - If the device is used transiently and results in changes to the anatomy and/or physiology that persist after use, the duration of device use is the terminal time point at which weight loss is measured. For example, for a device that is used temporarily but permanently reduces the size of the stomach, if the change in total body weight was assessed at 12 months post-device use, then the duration would be 12 months.
 - For devices that are used on a recurring basis, the duration of device use is the course of time the device is used before measuring the results. For example, for a device that is used daily, if the change in total body weight is assessed after eight months of daily use, then the duration would be eight months.

735

736 In a hypothetical example, a device was temporarily placed in the stomach. A clinical

737 investigation included two groups: a treatment group that had the device placed via an

738 endoscopic procedure; and a sham group for the control arm, which underwent an endoscopic

739 procedure, but no device was placed. After six months, devices were removed from the treatment

740 group and the change in weight was measured for both groups, so the duration of this device use

741 is six months. The results showed that at least half (50%) of the treatment group lost at least 5%

742 of their starting body weight. The results also showed that the treatment group lost more of their

743 starting body weight than the sham group did, with a superiority margin of 3% more weight lost.

744 Thus, the device successfully met co-primary effectiveness endpoints of 50% responder rate and

745 at least 2% TBWL over sham when measured at device removal 6 months post implant. Based

746 on the recommendations in Table 5, this weight loss would be considered “short-term limited

747 weight loss.”

748

749 FDA intends to use the weight loss indication categories (Table 5), the Adverse Event

750 classification (Table 2), and the Evaluation Matrices decision aid (Figure 1) to compare the

751 weight loss demonstrated with the adverse event classification profile as part of the benefit-risk

752 assessment of a weight loss device (Table 4).

- 753
- 754
- 755
- 756
- 757
- 758
- 759
- 760
- 761
- 762
- 763
- 764
- 765
1. There are four proposed Evaluation Matrices (numbered 1-4 in Figure 1). There is one Evaluation Matrix corresponding with each of the four weight loss indication categories described in Table 5. An Evaluation Matrix is selected for a device based on the amount of weight loss demonstrated in a clinical study and the duration of device use, consistent with Table 5.
 2. Within each Evaluation Matrix, there are five columns for the five grades of adverse events described in Table 2. For each grade of adverse event, if there is a patient in the clinical study with that adverse event, then a lettered cell is intended to be selected based on the percentage of patients who experienced that grade of adverse event. The letter of the cells is for reference purposes only.
 3. The shading of each cell indicates the possible consideration for the device based on the corresponding grade of adverse events (the column the cell is in). White indicates that the

Contains Nonbinding Recommendations

Draft – Not for Implementation

766 weight loss to adverse event profile appears favorable. Light gray shading indicates that
 767 the weight loss to adverse event profile is uncertain. Dark gray shading indicates that the
 768 weight loss to adverse event profile appears unfavorable.

769 4. The Evaluation Matrix for a specific device may include some combination of cells with
 770 different shading. The overall risk of the device depends on the cell of greatest risk; thus,
 771 the cell with the darkest shading suggests the outcome of the decision aid.


772 5. The outcome from the Evaluation Matrix is considered as part of the totality of the
 773 benefit-risk determination (Table 4).

774 6. The matrices are provided as a decision aid, which is only one part of FDA’s assessment
 775 when evaluating whether probable benefit outweighs probable risk for the device for its
 776 conditions of use.

777

	Weight loss to adverse event profile appears favorable
	Weight loss to adverse event profile is uncertain
	Weight loss to adverse event profile appears unfavorable

Reminder: The matrices are a decision aid, which is only one part of FDA’s assessment of whether benefit outweighs risk for the device for its conditions of use.



778

779

780 **Figure 1.** Evaluation Matrices for comparing weight loss indication categories (Table 5) and
 781 adverse events classification.⁶⁴

⁶⁴ Lettering within the matrices is included for reference purposes only. For example, the cell corresponding to a Grade IV adverse event occurring at a rate of more than 25% of the time is lettered “P.”

Contains Nonbinding Recommendations

Draft – Not for Implementation

782

783 In a hypothetical example, suppose that in a clinical investigation, a device successfully met co-
784 primary effectiveness endpoints of 50% responder rate and a superiority margin of 3% TBWL
785 over sham control when measured at device removal six months post implant. Based on Table 5,
786 the weight loss indication category would be “short-term limited weight loss,” so the device
787 would be evaluated via Evaluation Matrix 1 in Figure 1. In the assessment of the clinical study:

- 788 • 50% of patients had Grade I adverse events, which corresponds to the light gray cell
789 A in Matrix 1 of Figure 1;
- 790 • 3% of patients had Grade II adverse events, which corresponds to the white cell I in
791 Matrix 1 of Figure 1;
- 792 • 0% of patients had Grade III adverse events; and
- 793 • 1% of patients had Grade IV adverse events, which corresponds to the dark gray cell
794 S in Matrix 1 of Figure 1.

795

796 Overall, the risk of the device is characterized by the prevalence of greatest risk observed in the
797 study, i.e., the 1% Grade IV adverse event rate, where the dark gray cell indicates that the weight
798 loss to adverse event profile may not be favorable for the given amount of weight loss as part of
799 the overall benefit-risk assessment. The low rate of adverse events in Grade II (the white cell)
800 and Grade I (the light gray cell) may not negate the risk associated with the rate of adverse
801 events in Grade IV (the dark gray cell).

802

803 In another hypothetical example, suppose that in a clinical investigation, a device successfully
804 met co-primary effectiveness endpoints of 50% responder rate and a superiority margin of 10%
805 TBWL over sham control when measured at device removal 12 months post implant. Based on
806 Table 5, the weight loss indication category would be “weight loss,” so the device would be
807 evaluated via Evaluation Matrix 4 in Figure 1. In the assessment of the clinical study:

- 808 • 70% of patients had Grade I adverse events, which corresponds to the light gray cell
809 A in Matrix 4 of Figure 1;
- 810 • 10% of patients had Grade II adverse events, which corresponds to the white cell G in
811 Matrix 4 of Figure 1;
- 812 • 0.5% of patients had Grade III adverse events, which corresponds to the white cell O
813 in Matrix 4 of Figure 1; and
- 814 • 2% of patients had Grade IV adverse events, which corresponds to the light gray cell
815 S in Matrix 4 of Figure 1.

816

817 Overall, the risk of the device is characterized by the prevalence of greatest risk observed in the
818 study, i.e., the 2% Grade IV adverse event rate and 70% Grade I adverse event rate, where the
819 light gray cells indicate that the weight loss to adverse event profile is uncertain given the
820 amount of weight loss as part of the overall benefit-risk assessment. The low rate of adverse
821 events in Grade II and Grade III (the white cells) may not negate the risk associated with the rate
822 of adverse events in Grade I and Grade IV (the light gray cells).

823

Contains Nonbinding Recommendations

Draft – Not for Implementation

824 During the review of a marketing submission, FDA intends to consider information from the
825 proposed Evaluation Matrices, along with all other applicable factors identified in Table 4, to
826 make a final determination regarding whether the probable benefits of the device outweigh the
827 probable risks of the device.

DRAFT